


SOLUTION CHANGE

DataMaster DMT: 103709
Location: St. Johnsbury VSP
Date: 09/03/2016
Performed by: TYLOR D RANCOURT



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-24-100
Range = 0.094 - 0.104
Average = 0.100 g/210L
Std Dev = 0.0004

ROD 9/3/16

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read 'T. Rancourt', is written over a horizontal line.

Date

09/03/2016